Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Product Name : Bottle m FC Agar 500G
Catalog Number : 267720 **Manufacture Date**: 2024/02/01
Batch Number : 4073947
Expiration Date : 2028/01/31

01. Dehydrated Medium Appearance: Beige with slight blue tint to blue, free-flowing, homogeneous
02. Solubility: 5.2% solution, soluble in distilled or deionized water on boiling
03. Solution Appearance: Plain: medium to dark blue, very slightly to slightly opalescent, may have slight precipitate.
With addition of 1% Rosolic Acid: purple to reddish purple to cranberry red, slightly opalescent, may have slight precipitate
04. Cultural Response: Medium was prepared per label instructions with 1% Rosolic Acid. Plates were inoculated using the membrane filter technique and incubated at 44.5 ± 0.5°C for 24 ± 2 hours.

TEST ORGANISMS	ATCC®	RECOVERY	COLOR OF COLONY
Enterococcus faecalis	19433	markedly to completely inhibited	- - -
Escherichia coli	25922	good	blue

Characteristic	Unit	Value	Lower Limit	Upper Limit
pH at 25°C :		7.3	7.2	7.6
Bulk Lot Number:	-	4029866		

Animal source	Country of Origin	Tissue Category BIC	SIC	ABC
Bovine	Argentina	IV	IV	IC
Bovine	Mexico	IV	IV	IC
Bovine	New Zealand	IV	IV	IC
Bovine	New Zealand	IV	IV	MLK
Bovine	USA	IV	IV	IC
Bovine	USA	IV	IV	MLK
Ovine	New Zealand	IV	IV	IC
Porcine	USA	II	II	IB
Porcine	USA	III	III	IB

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

BD Diagnostics - Diagnostic Systems products are manufactured in ISO

Creation Date: 2024/04/19 14:45:19

Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Product Name : Bottle m FC Agar 500G
Catalog Number : 267720 **Manufacture Date**: 2024/02/01
Batch Number : 4073947
Expiration Date : 2028/01/31

13485:2016 Registered facilities. In addition, BD Diagnostics - Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at www.bd.com/regdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.

For complete details on animal origin information, refer to "BD Diagnostics - Diagnostic Systems, Animal Origin Position Statement", at www.bd.com/regdocs under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/en-us/support/bd-life-sciences-diagnostic-systems-customer-regulatory-support-information.

Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Product Name : Bottle m FC Agar 500G
Catalog Number : 267720 **Manufacture Date**: 2024/02/01
Batch Number : 4073947
Expiration Date : 2028/01/31



Stacie Wright
Associate Director, Quality Assurance
Signature Date: 2024/04/19

Creation Date: 2024/04/19 14:45:19