Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Product Name : Bottle M Endo Agar LES 500G
Catalog Number : 273620 **Manufacture Date**: 2023/02/23
Batch Number : 3300597
Expiration Date : 2028/01/31

01. Dehydrated Medium Appearance: Purple, free-flowing, homogeneous
02. Solubility: 5.1% solution, soluble in distilled or deionized water containing 2% ethanol on boiling
03. Solution Appearance: Pinkish-red, slightly opalescent to opalescent with precipitation
04. Cultural Response: Medium was prepared per label instructions. Using the membrane filter technique, filters were inoculated with the test organisms and preincubated for 1½-2 hours at 35 ± 0.5°C on pads saturated with Lauryl Tryptose Broth. Filters were then transferred to plates with m Endo Agar LES and incubated at 35 ± 0.5°C for 22 ± 2 hours.

TEST ORGANISMS	ATCC®	RECOVERY	COLONY COLOR
Escherichia coli	25922	good	red with sheen
Salmonella typhimurium	14028	good	colorless to pink
Staphylococcus aureus	25923	marked to complete inhibition	- - - -

Characteristic	Unit	Value	Lower Limit	Upper Limit
pH at 25°C :		7.3	7.0	7.4
Bulk Lot Number :	-	3045535		

Animal source	Country of Origin	Tissue Category BIC	SIC	ABC
Bovine	Australia	IV	IV	IC
Bovine	Brazil	IV	IV	IC
Bovine	Mexico	IV	IV	IC
Bovine	New Zealand	IV	IV	IC
Bovine	New Zealand	IV	IV	MLK
Bovine	USA	IV	IV	IC
Bovine	USA	IV	IV	MLK
Ovine	New Zealand	IV	IV	IC
Porcine	Canada	IV	IV	IB
Porcine	USA	II	II	IB
Porcine	USA	III	III	IB
Porcine	USA	IV	IV	IB

The Batch Number on this certificate is synonymous with the Lot Number shown on the product label.

Creation Date: 2023/12/20 16:14:59

Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Product Name : Bottle M Endo Agar LES 500G
Catalog Number : 273620 **Manufacture Date**: 2023/02/23
Batch Number : 3300597
Expiration Date : 2028/01/31

BD Diagnostics - Diagnostic Systems products are manufactured in ISO 13485:2016 Registered facilities. In addition, BD Diagnostics - Diagnostic Systems facilities are registered with the United States Food and Drug Administration (FDA), and are regulated by the FDA's Quality System Regulations (QSRs). This product met BDDS stringent quality standards at time of batch/lot release. Any test results reported on this certificate were obtained at time of release. This material is not for human or animal consumption.

BD Diagnostics - Diagnostic Systems' Certificates of Analysis (COA) typically contain animal origin information when products are manufactured using materials of animal origin. This information may be contained in the animal source table and/or in one or more of the additional paragraphs found on the COA. Following Quality Control release, the COA is created and published at www.bd.com/regdocs. For each batch of finished product that contains animal origin raw materials, the COA shows the animal origin data from the individual lots of animal origin raw materials used, as provided by the raw material suppliers.

At times, suppliers notify BD Diagnostics - Diagnostic Systems of new and/or additional information they have received from their raw material suppliers that modifies the animal origin information for lots previously provided to BD. See "COA Animal Origin Information Position Statement" located at www.bd.com/regdocs under "Position Statements" for the impact that retrospective information has on COAs and on customers enrolled in the BDDS and BDAB Automated Change Notification Program.


For complete details on animal origin information, refer to "BD Diagnostics - Diagnostic Systems, Animal Origin Position Statement", at www.bd.com/regdocs under "Position Statements".

Manufacturer is Becton Dickinson and Company, 7 Loveton Circle, Sparks, MD 21152 USA. To determine location of manufacturing for this product, please see www.bd.com/en-us/support/bd-life-sciences-diagnostic-systems-customer-regulatory-support-information.

Certificate of Analysis

Becton Dickinson and Company
BD Diagnostic Systems
PO Box 999
Sparks MD 21152-0999 US

Product Name : Bottle M Endo Agar LES 500G
Catalog Number : 273620 **Manufacture Date**: 2023/02/23
Batch Number : 3300597
Expiration Date : 2028/01/31



Zuleika Vargas
Director, Quality Management
Signature Date:2023/12/20